



Ref. Certif. No.

SG-MD-00248

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC**Product
ProduitGeneral Medical Devices
Alternating Pressure Mattress With PumpName and address of the applicant
Nom et adresse du demandeurXiaMen SenYang Co., Ltd
4-5 Floor, Xingbei Industry, NO.95-99, West 2 Road, Jiu Tianhu, Xinglin
361000 Xiamen, PEOPLE'S REPUBLIC OF CHINAName and address of the manufacturer
Nom et adresse du fabricantXiaMen SenYang Co., Ltd, 4-5 Floor, Xingbei Industry, NO.95-99, West 2
Road, Jiu Tianhu, Xinglin, 361000 Xiamen, PEOPLE'S REPUBLIC OF
CHINAName and address of the factory
Nom et adresse de l'usineXiaMen SenYang Co., Ltd, 4-5 Floor, Xingbei Industry, NO.95-99, West 2
Road, Jiu Tianhu, Xinglin, 361000 Xiamen, PEOPLE'S REPUBLIC OF
CHINARatings and principal characteristics
Valeurs nominales et caractéristiques principalesRated voltage: 220V-240V~
Rated frequency: 50Hz
Rated current: 50mA
Protection class: II
Applied part: Type BModel/type Ref.
Ref. de type


APP-P05 , APP-P01

Additional information (if necessary)
Information complémentaire (si nécessaire)

See general product information in test report

A sample of the product was tested and found
to be in conformity with

IEC 60601-1:2005

Un échantillon de ce produit a été essayé et a été
considéré conforme à laas shown in the Test Report Ref. No.
which forms part of this certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue une partie de ce
certificatTÜV SÜD PSB Pte Ltd
64.66T.11.039.01This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de CertificationDate, 2012-06-27
CBS 12 06 78002 002
(Ben Xu)

TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221

PSB Singapore